
	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ3U00AZ.01
---	---	-------------

TECHNOVIEW Arixtra CONT L 6x1ml

REF 5090012
LOT 3U31B00.01
 30-APR-2015

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.4 – 7.8	7.5	OK
Virologie Virology	HbsAg HIV 1,2 Ab HCV	Negativ Negative	Negativ Negative	OK
Richtwert Expected value Ceveron® alpha	Anti Xa (µg/L)	0.41 – 0.75µg/mL	0.58 µg/mL	
Verwendbarkeit Usability Gerät	Anti Xa (3U31B00)	0.41 – 0.75µg/mL	0.63µg/mL	OK

**Durchführung von
Prepared by**

Lucas HOŁOVICZ

16-MAI-2013
Datum / Date:

**Freigabe von
Approved by**

Harald SEVCIK

16. MAI 2013
Datum / Date: