



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQ3U00AZ.01

TECHNOVIEW Arixtra Control L 6x1mL

REF 5090012
LOT 3U31B01.02
⌚ 30-Sep-2016

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.4 – 7.8	7.5	OK
Virologie Virology	HbsAg HIV 1,2 Ab HCV	Negativ Negativ Negativ	Negativ Negativ Negativ	OK
Richtwert Expected value Ceveron® alpha	Anti Xa	0.41 – 0.75 µg/mL	0.58	
Verwendbarkeit Usability Ceveron® alpha	Anti Xa	0.41 – 0.75 µg/mL	0.51	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

Durchführung von
Prepared by


Lucas ROSENMAYR

09-Jun-2015
Datum / Date:

Freigabe von
Approved by


Lavdrim NASUFI

09-Jun-2015
Datum / Date: