
	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ3V00AZ.01
---	---	-------------

TECHNOVIEW Arixtra CONT H 6x1ml

REF 5090014
LOT 3V31B00.01
 30-APR-2015

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.4 – 7.8	7.6	OK
Virologie Virology	HbsAg HIV 1,2 Ab HCV	Negativ Negative	Negativ Negative	OK
Richtwert Expected value Ceveron® alpha	Anti Xa (µg/mL)	0.71 – 1.07µg/mL	0.89µg/mL	
Verwendbarkeit Usability Gerät	Anti Xa (3U31B00)	0.71 – 1.07µg/mL	0.83µg/mL	OK

**Durchführung von
Prepared by**

Lucas HOLOVICZ



16-MAI-2013
Datum / Date:

**Freigabe von
Approved by**

Harald SEVCIK



16. MAI 2013
Datum / Date: