



ANALYSENZERTIFIKAT  
CERTIFICATE OF ANALYSIS

DQ3V00AZ.01

TECHNOVIEW Arixtra CONT H 6x1ml

REF 5090014  
LOT 3V31B01.01  
⌚ 30-Sep-2016

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.4 – 7.8	7.6	OK
Virologie Virology	HbsAg HIV 1,2 Ab HCV	Negativ Negative	Negativ Negative	OK
Richtwert Expected value Ceveron® alpha	Anti Xa	0.71 – 1.07µg/mL	0.89	
Verwendbarkeit Usability Ceveron® alpha	Anti Xa	0.71 – 1.07µg/mL	0.98	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

Durchführung von  
Prepared by

Lucas HOLOVICZ

03-Oct-2014  
Datum / Date:

Freigabe von  
Approved by

Alessandro BENIGNI

03-Oct-2014  
Datum / Date: