



| | | |
|---|---|-------------|
|  | ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS | DQ3V00AZ.01 |
|---|---|-------------|

TECHNOVIEW Arixtra Control H 6x1mL

REF 5090014
LOT 3V31B01.02
 30-Sep-2016

| Analytische Prozedur Analytical Procedure | | Anforderung Requirement | Ergebnis Result | |
|--|----------------------------|-------------------------------|-------------------------------|----|
| pH | | 7.4 – 7.8 | 7.6 | OK |
| Virologie Virology | HbsAg HIV 1,2 Ab HCV | Negativ Negativ Negativ | Negativ Negativ Negativ | OK |
| Richtwert Expected value Ceveron® alpha | Anti Xa | 0.71 – 1.07 µg/mL | 0.89 | |
| Verwendbarkeit Usability Ceveron® alpha | Anti Xa | 0.71 – 1.07 µg/mL | 0.83 | OK |
| Etiketten und Barcode Überprüfung Label and bar code Verification | | | OK | |

**Durchführung von
Prepared by**

Lucas ROSENMAYR

09-Jun-2015
Datum / Date:

**Freigabe von
Approved by**

Lavdrim NASUFI

09-Jun-2015
Datum / Date: