
	<b>ANALYSENZERTIFIKAT</b> <b>CERTIFICATE OF ANALYSIS</b>	DQ3V00AZ.01
---	---	-------------

### TECHNOVIEW Arixtra CONT H 6x1mL

REF      5090014  
LOT      3V51B00.01  
      31-Aug-2017


Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.4 – 7.8	7.7	OK
Virologie Virology	HBsAg / HBsAg HIV-AK / HIV-Ab HCV-AK / HCV-Ab	Negative Negative Negative	Negative Negative Negative	OK
Richtwert Expected value Ceveron® alpha	Anti Xa	0.62 – 1.04 µg/mL	0.83	
Verwendbarkeit Usability Ceveron® alpha	Anti Xa	0.62 – 1.04 µg/mL	0.87	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

**Durchführung von**  
**Prepared by**

  
**Lucas ROSENMAYR**

14-Oct-2015  
Datum / Date:

**Freigabe von**  
**Approved by**

  
**Lavdrim NASUFI**

14-Oct-2015  
Datum / Date: