



**ANALYSENZERTIFIKAT  
CERTIFICATE OF ANALYSIS**

DQ0H00AZ.01

**TECHNOVIEW LMW CAL Set 5x1mL**

**REF** 5090040

**LOT** 0H31B00.01

31-JUL-2015

Inhalt Content	Reagenz Reagent	<b>LOT</b>	
1x1mL	TECHNOVIEW LMW Cal 1	5A15B01	31-JUL-2015
1x1mL	TECHNOVIEW LMW Cal 2	5B12B01	31-JUL-2015
1x1mL	TECHNOVIEW LMW Cal 3	5C12B01	31-JUL-2015
1x1mL	TECHNOVIEW LMW Cal 4	5D12B01	31-JUL-2015
1x1mL	TECHNOVIEW LMW Cal 5	5E13B01	31-JUL-2015

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Verwendbarkeit Usability Ceveron <sup>®</sup> alpha	TECHNOVIEW LMW Cont low (3F21B00)	0.29 – 0.53 IU/mL	0.45 IU/mL	OK
	TECHNOVIEW LMW Cont high (3H12B00)	1.01 – 1.68 IU/mL	1.34 IU/mL	OK

Durchführung von  
Prepared by

**Lucas HOLOVICZ**

16. AUG. 2013  
Datum / Date:

Freigabe von  
Approved by

**Harald SEVCIK**

16. AUG. 2013  
Datum / Date: