

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ0Y00AZ.01
---	---	-------------

TECHNOVIEW UFH CAL Set 5x1mL

REF 5090070
LOT 0Y31B00.01
 31-JUL-2015

Inhalt Content	Reagenz Reagent	LOT	
1x1mL	TECHNOVIEW UFH Cal 1	5F13B01	31-JUL-2015
1x1mL	TECHNOVIEW UFH Cal 2	5G13B01	31-JUL-2015
1x1mL	TECHNOVIEW UFH Cal 3	5H13B01	31-JUL-2015
1x1mL	TECHNOVIEW UFH Cal 4	5J13B01	31-JUL-2015
1x1mL	TECHNOVIEW UFH Cal 5	5K13B01	31-JUL-2015

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Verwendbarkeit Usability Ceveron [®] alpha	TECHNOVIEW UFH Cont low (3C12B00)	0.24 – 0.44 IU/mL	0.27 IU/mL	OK
	TECHNOVIEW UFH Cont high (3J12B00)	0.77 – 1.29 IU/mL	0.87 IU/mL	OK

Durchführung von
Prepared by

Lucas HOŁOVIĆZ

13. Aug - 2013
Datum / Date:

Freigabe von
Approved by

Harald SEVCIK

13. AUG. 2013
Datum / Date: