



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQ3300AZ.01

Technoview Rivaroxaban CAL High 5x1mL

REF 5090171
LOT 0331B00.01
⌚ 31-Jul-2015

Inhalt Content	Reagenz Reagent	LOT	⌚
1X1mL	Technoview Rivaroxaban CAL1 High	5101B01	31-Dez-2016
1X1mL	Technoview Rivaroxaban CAL2 High	5201B01	31-Dez-2016
1X1mL	Technoview Rivaroxaban CAL3 High	5302B01	31-Dez-2016
1X1mL	Technoview Rivaroxaban CAL4 High	5411B01	31-Jul-2015
1X1mL	Technoview Rivaroxaban CAL5 High	5501B01	31-Dez-2016

Analytische Prozedur Analytical Procedure	Anforderung Requirement	Ergebnis Result		
Verwendbarkeit Usability Ceveron® alpha	Rivaroxaban Control High LOT 3301B00	244.2 - 366.4 ng/mL	329.4 ng/mL	OK
	Rivaroxaban Control Low LOT 3402B01	42.0 - 78.0 ng/mL	61.9 ng/mL	OK

Durchführung von
Prepared by

Alessandro BENIGNI

24-Jul-2013
Datum / Date:

Freigabe von
Approved by

Harald SEVCIK

24. JULI 2013
Datum / Date: