

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ3300AZ.01
---	---	-------------

Technoview Rivaroxaban CAL High 5x1mL

REF 5090171
LOT 0331B00.03
 31-Jul-2015

Inhalt Content	Reagenz Reagent	LOT	
1X1mL	Technoview Rivaroxaban CAL1 High	5101B01	31-Dez-2016
1X1mL	Technoview Rivaroxaban CAL2 High	5201B01	31-Dez-2016
1X1mL	Technoview Rivaroxaban CAL3 High	5302B01	31-Dez-2016
1X1mL	Technoview Rivaroxaban CAL4 High	5411B01	31-Jul-2015
1X1mL	Technoview Rivaroxaban CAL5 High	5501B01	31-Dez-2016

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Verwendbarkeit Usability Ceveron [®] alpha	Rivaroxaban Control High LOT 3331B00	252.5 – 378.7 ng/mL	318.5	OK
	Rivaroxaban Control Med LOT 3532B00	108.2 – 180.4 ng/mL	159.6	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

Durchführung von
Prepared by

Lucas HOLOVICZ

10-Nov-2014
Datum / Date:

Freigabe von
Approved by

Fikret NASUFI

10-Nov-2014
Datum / Date: