

	<b>ANALYSENZERTIFIKAT</b> <b>CERTIFICATE OF ANALYSIS</b>	DQ0300.01
---	---	-----------

### Technoview Rivaroxaban CAL High 5x1mL

REF      5090171  
LOT      0351B00.01  
      31-Dez-2016

Inhalt Content	Reagenz Reagent	<span style="border: 1px solid black; padding: 2px;">LOT</span>	
1X1mL	Technoview Rivaroxaban CAL1 High	5101B01	31-Dez-2016
1X1mL	Technoview Rivaroxaban CAL2 High	5201B01	31-Dez-2016
1X1mL	Technoview Rivaroxaban CAL3 High	5302B01	31-Dez-2016
1X1mL	Technoview Rivaroxaban CAL4 High	5411B02	31-Jan-2017
1X1mL	Technoview Rivaroxaban CAL5 High	5501B01	31-Dez-2016

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Verwendbarkeit Usability Ceveron <sup>®</sup> alpha	Rivaroxaban Control High <span style="border: 1px solid black; padding: 2px;">LOT</span> 3301B01	244.2 – 366.4 ng/mL	333.2	OK
	Rivaroxaban Control Low <span style="border: 1px solid black; padding: 2px;">LOT</span> 3431B01		51.0 – 76.6 ng/mL	66.9
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

**Durchführung von**  
**Prepared by**

Lucas Rosemayr

17-Feb-2015  
Datum / Date:

**Freigabe von**  
**Approved by**

Lucas HOLOVICZ

17-Feb-2015  
Datum / Date: