

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ0300.01
---	---	-----------

Technoview® Rivaroxaban CAL Set High 5x1mL

REF 5090171
LOT 0351B00.02
 31-Dez-2016

Inhalt Content	Reagenz Reagent	LOT	
1X1mL	Technoview Rivaroxaban CAL1 High	5101B01	31-Dez-2016
1X1mL	Technoview Rivaroxaban CAL2 High	5201B01	31-Dez-2016
1X1mL	Technoview Rivaroxaban CAL3 High	5302B01	31-Dez-2016
1X1mL	Technoview Rivaroxaban CAL4 High	5411B02	31-Jan-2017
1X1mL	Technoview Rivaroxaban CAL5 High	5501B01	31-Dez-2016

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Verwendbarkeit Usability Ceveron® alpha	Rivaroxaban Control High LOT 3331B01	252.5 – 378.7 ng/mL	326.9	OK
	Rivaroxaban Control Low LOT 3451B01	53.9 – 80.9 ng/mL	61.7	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

Durchführung von
Prepared by


Monika WEINBAUER

12. OKT. 2015

Datum / Date:

Freigabe von
Approved by


Laydrim NASUFI

13. Okt. 2015

Datum / Date: