
	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ3300AZ.01
---	---	-------------

TECHNOVIEW® Rivaroxaban Control Low

REF 5090172
LOT 3431B00.02
 31-Mai-2015

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.4 – 7.8	7.6	OK
Virologie Virology	HbsAg HIV 1,2 Ab HCV	Negativ Negative	Negativ Negative	OK
Chargenspezifische Kalibrationswerte Lot specific calibration values		51.0 – 76.6 ng/ml	63.8	OK
Verwendbarkeit Usability Ceveron® alpha	TECHNOCHROM® anti-Xa (ng/mL)	51.0 – 76.6 ng/ml	57.5	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

**Durchführung von
Prepared by**

Lucas HOLOVICZ

26-Aug-2014
Datum / Date:

**Freigabe von
Approved by**

Michael RIHA

26-Aug-2014
Datum / Date: