
	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ3400.01
---	---	-----------

TECHNOVIEW Rivaroxaban Control low 5x1ml

REF 5090172
LOT 3431B01.01
 31-Jan-2017

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.4 – 7.8	7.6	OK
Virologie Virology	HbsAg HIV 1,2 Ab HCV	Negativ Negative	Negativ Negative	OK
Chargenspezifische Kalibrationswerte Lot specific calibration values		51.0 – 76.6 ng/mL	63.8	OK
Verwendbarkeit Usability Gerät	TECHNOCHROM® anti-Xa (ng/mL)	51.0 – 76.6 ng/mL	66.9	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

Durchführung von
Prepared by

Lucas Rosenmayr

17-Feb-2015
Datum / Date:

Freigabe von
Approved by

Lucas HOLOVICZ

17-Feb-2015
Datum / Date: