
	<b>ANALYSENZERTIFIKAT</b> <b>CERTIFICATE OF ANALYSIS</b>	DQ3400.01
---	---	-----------

**TECHNOVIEW Rivaroxaban Control low 5x1mL**

REF      5090172  
LOT      3451B00.01  
      28-Feb-2019

<b>Analytische Prozedur</b> <b>Analytical Procedure</b>		<b>Anforderung</b> <b>Requirement</b>	<b>Ergebnis</b> <b>Result</b>	
pH		7.4 – 7.8	7.7	OK
Virologie Virology	HbsAg HIV 1,2 Ab HCV	Negative Negative Negative	Negative Negative Negative	OK
Chargenspezifische Kalibrationswerte Lot specific calibration values		53.9 – 80.9 ng/mL	67.4	OK
Verwendbarkeit Usability Gerät	TECHNOCHROM® anti-Xa (ng/mL)	53.9 – 80.9 ng/mL	61.0	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

**Durchführung von**  
**Prepared by**

**Lucas ROSENMAYR**

*28-Apr-2015*  
Datum / Date:

**Freigabe von**  
**Approved by**

**Lavdrim NASUFI**

*28-Apr-2015*  
Datum / Date: