
	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ3500AZ.01
---	---	-------------

TECHNOVIEW Rivaroxaban CONT M 5x1ml

REF 5090173
LOT 3532B01.01
 30-Nov-2017

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Analytische Verfahren Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.4 – 7.8	7.6	OK
Virologie Virology	HBsAg / HBsAg HIV-AK / HIV-Ab HCV-AK / HCV-Ab	Negative Negative Negative	Negative Negative Negative	OK

Verwendbarkeit Usability Gerät Ceveron Alpha®	Anti Xa Rivaroxaban High	108.2 – 180.4 ng/mL	164.2	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

**Durchführung von
Prepared by**

Lucas HOLOVICZ

12-Jan-2015
Datum / Date:

**Freigabe von
Approved by**

Fikret NASUFI

12-Jan-2015
Datum / Date: