
	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ3300AZ.01
---	---	-------------

TECHNOVIEW® Rivaroxaban Control High

REF 5090174
LOT 3331B00.03
 30-Jun-2015

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.4 – 7.8	7.5	OK
Virologie Virology	HbsAg HIV 1,2 Ab HCV	Negativ Negative	Negativ Negative	OK
Chargenspezifische Kalibrationswerte Lot specific calibration values		252.5 – 378.7 ng/mL	315.6	
Verwendbarkeit Usability Ceveron® alpha	TECHNOCHROM® anti-Xa (ng/mL)	252.5 – 378.7 ng/mL	307.0	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

Durchführung von
Prepared by

Lucas HOLOVICZ

16-Jun-2014
Datum / Date:

Freigabe von
Approved by

Lavdrim NASUFI

16-Jun-2014
Datum / Date: