
	<b>ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS</b>	DQ3300AZ.01
---	---	-------------

## TECHNOVIEW® Rivaroxaban Control High

REF      5090174  
LOT      3331B00.04  
      30-Jun-2015

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.4 – 7.8	7.5	OK
Virologie Virology	HbsAg HIV 1,2 Ab HCV	Negativ Negative	Negativ Negative	OK
Chargenspezifische Kalibrationswerte Lot specific calibration values		252.5 – 378.7 ng/mL	315.6	
Verwendbarkeit Usability Ceveron® alpha	TECHNOCHROM® anti-Xa (ng/mL)	252.5 – 378.7 ng/mL	324.4	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

**Durchführung von  
Prepared by**

Lucas HOLÓVICZ

26-JUN-2015  
Datum / Date:

**Freigabe von  
Approved by**

Michael RIHA

26-JUN-2015  
Datum / Date: