
	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ3300.01
---	---	-----------

TECHNOVIEW® Rivaroxaban Control High 5x1mL

REF 5090174
LOT 3331B01.01
 30-Nov-2017

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.4 – 7.8	7.5	OK
Virologie Virology	HbsAg HIV 1,2 Ab HCV	Negativ Negative	Negativ Negative	OK
Chargenspezifische Kalibrationswerte Lot specific calibration values		252.5 – 378.7 ng/mL	315.6	
Verwendbarkeit Usability Ceveron® alpha	TECHNOCHROM® anti-Xa (ng/mL)	252.5 – 378.7 ng/mL	324.4	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

Durchführung von
Prepared by

Lucas Rosenmayr

17-Feb-2015
Datum / Date:

Freigabe von
Approved by

Lucas HOLOVICZ

17-Feb-2015
Datum / Date: