




ANALYSENZERTIFIKAT  
CERTIFICATE OF ANALYSIS

DQ3700.01

TECHNOVIEW Dabigatran CONT Low 5x1ml

REF 5090214


LOT 3742B00.01

 29-Feb-2016

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.4 – 7.8	7.6	OK
Virologie Virology	HBsAg / HBsAg HIV-AK / HIV-Ab HCV-AK / HCV-Ab	Negative Negative Negative	Negative Negative Negative	OK
Chargenspezifische Kalibrationswerte Lot specific calibration values		98.9 – 183.7 ng/mL	141.3	OK
Verwendbarkeit Usability Ceveron ® alpha	Anti-Xa Dabigatran 3742B00	98.9 – 183.7 ng/mL	128.2	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	


Durchführung von  
Prepared by

  
Lucas HOLOVIC

  
Datum / Date:

Freigabe von  
Approved by

  
Alessandro BENIGNI

  
Datum / Date: