

	<b>ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS</b>	DQ1000.01
---	---	-----------

### TECHNOVIEW Apixaban CAL 5x1mL

REF      5090269  
LOT      1061B00.01  
      31-Jul-2017

Inhalt Content	Reagenz Reagent	<span style="border: 1px solid black; padding: 2px;">LOT</span>	
1x1mL	TECHNOVIEW Apixaban Cal 1	1142B01	31-Jul-2017
1x1mL	TECHNOVIEW Apixaban Cal 2	1243B01	31-Jul-2017
1x1mL	TECHNOVIEW Apixaban Cal 3	1341B01	31-Jul-2017
1x1mL	TECHNOVIEW Apixaban Cal 4	1443B01	31-Jul-2017
1x1mL	TECHNOVIEW Apixaban Cal 5	1543B01	31-Jul-2017

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Verwendbarkeit Usability Ceveron® alpha	TECHNOVIEW Apixaban Cont low <span style="border: 1px solid black; padding: 2px;">LOT</span> 3941B01	104.2 – 173.6 ng/mL	145.3	OK
	TECHNOVIEW Apixaban Cont high <span style="border: 1px solid black; padding: 2px;">LOT</span> 3D43B01	282.3 – 470.5 ng/mL	377.7	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

**Durchführung von  
Prepared by**

  
**Monika WEINBAUER**

08. MRZ. 2016

Datum / Date:

**Freigabe von  
Approved by**

  
**Lavdrim NASUFI**

08. MRZ. 2016

Datum / Date: