

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ1000.01
---	---	-----------


TECHNOVIEW Apixaban CAL 5x1mL

REF 5090269
LOT 1032B00.01
 31-Dec-2015

Inhalt Content	Reagenz Reagent	LOT	
1x1mL	TECHNOVIEW Apixaban Cal 1	1132B00	31-Dec-2015
1x1mL	TECHNOVIEW Apixaban Cal 2	1232B00	31-Dec-2015
1x1mL	TECHNOVIEW Apixaban Cal 3	1332B00	31-Dec-2015
1x1mL	TECHNOVIEW Apixaban Cal 4	1432B00	31-Dec-2015
1x1mL	TECHNOVIEW Apixaban Cal 5	1532B00	31-Dec-2015


Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Verwendbarkeit Usability Ceveron [®] alpha	TECHNOVIEW Apixaban Cont low (3932B00)	98.9 – 164.8 ng/mL	122.0	OK
	TECHNOVIEW Apixaban Cont high (3D32B00)	253.5 – 422.5 ng/mL	347.1	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

Durchführung von
Prepared by

Lucas HOLOVICZ


25-Feb-2014
Datum / Date:

Freigabe von
Approved by

Lavdrim NASUFI


25-Feb-2014
Datum / Date: