

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ1000.01
---	---	-----------

TECHNOVIEW Apixaban CAL 5x1mL


REF 5090269
LOT 1041B00.01
 29-Feb-2016

Inhalt Content	Reagenz Reagent	LOT	
1x1mL	TECHNOVIEW Apixaban Cal 1	1142B00	29-Feb-2016
1x1mL	TECHNOVIEW Apixaban Cal 2	1242B00	29-Feb-2016
1x1mL	TECHNOVIEW Apixaban Cal 3	1341B00	29-Feb-2016
1x1mL	TECHNOVIEW Apixaban Cal 4	1442B00	29-Feb-2016
1x1mL	TECHNOVIEW Apixaban Cal 5	1542B00	29-Feb-2016

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Verwendbarkeit Usability Ceveron [®] alpha	TECHNOVIEW Apixaban Cont low (3941B00)	104.2 – 173.6 ng/mL	152.3	OK
	TECHNOVIEW Apixaban Cont high (3D41B00)	261.2 – 435.4 ng/mL	349.7	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	


**Durchführung von
Prepared by**


Lucas HOLOVICZ


Datum / Date:

**Freigabe von
Approved by**


Alessandro BENIGNI


Datum / Date: