





**ANALYSENZERTIFIKAT  
CERTIFICATE OF ANALYSIS**

DQ1000.01

**TECHNOVIEW Apixaban CAL 5x1mL**

**REF**            5090269  
**LOT**            1041B01.01  
            31-Jul-2017

Inhalt Content	Reagenz Reagent	<b>LOT</b>	
1x1mL	TECHNOVIEW Apixaban Cal 1	1142B01	31-Jul-2017
1x1mL	TECHNOVIEW Apixaban Cal 2	1242B01	31-Jul-2017
1x1mL	TECHNOVIEW Apixaban Cal 3	1341B01	31-Jul-2017
1x1mL	TECHNOVIEW Apixaban Cal 4	1442B01	31-Jul-2017
1x1mL	TECHNOVIEW Apixaban Cal 5	1542B01	31-Jul-2017

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Verwendbarkeit Usability Ceveron <sup>®</sup> alpha	TECHNOVIEW Apixaban Cont low <b>LOT</b> 3941B01	104.2 – 173.6 ng/mL	135.4	OK
	TECHNOVIEW Apixaban Cont high <b>LOT</b> 3D41B01	261.2 – 435.4 ng/mL	396.2	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

**Durchführung von  
Prepared by**

Lucas ROSENMAYP

26-Aug-2015  
Datum / Date:

**Freigabe von  
Approved by**

Lucas HOLOVICZ

26-Aug-2015  
Datum / Date: