

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ1000.01
---	---	-----------

TECHNOVIEW Apixaban CAL 5x1mL

REF 5090269
LOT 1042B01.01
 31-Jul-2017

Inhalt Content	Reagenz Reagent	LOT	
1x1mL	TECHNOVIEW Apixaban Cal 1	1143B01	31-Jul-2017
1x1mL	TECHNOVIEW Apixaban Cal 2	1243B01	31-Jul-2017
1x1mL	TECHNOVIEW Apixaban Cal 3	1343B01	31-Jul-2017
1x1mL	TECHNOVIEW Apixaban Cal 4	1443B01	31-Jul-2017
1x1mL	TECHNOVIEW Apixaban Cal 5	1543B01	31-Jul-2017

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Verwendbarkeit Usability Ceveron [®] alpha	TECHNOVIEW Apixaban Cont low LOT 3941B01	104.2 – 173.6 ng/mL	138.8	OK
	TECHNOVIEW Apixaban Cont high LOT 3D41B01	261.2 – 435.4 ng/mL	304.3	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

Durchführung von
Prepared by


Lucas ROSENMAYR

26-Aug-2015
Datum / Date:

Freigabe von
Approved by


Monika WEINBAUER

26-Aug-2015
Datum / Date: