
	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ3D00.01
---	---	-----------

TECHNOVIEW Apixaban CONT High 5x1ml

REF 5090270
LOT 3D32B00.01
 31-Dec-2015

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.4 – 7.8	7.6	OK
Virologie Virology	HBsAg / HBsAg HIV-AK / HIV-Ab HCV-AK / HCV-Ab	Negative Negative Negative	Negative Negative Negative	OK
Chargenspezifische Kalibrationswerte Lot specific calibration values		2.30 – 3.84U/mL	3.07	OK
		253.5 – 422.5ng/mL	338.0	
Verwendbarkeit Usability Ceveron ® alpha	Anti-Xa Apixaban 3931B00	253.5 – 422.5ng/mL	347.1	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

**Durchführung von
Prepared by**

Lucas HOLOVICZ

25-Feb-2014
Datum / Date:

**Freigabe von
Approved by**

Laydrim NASUFI

25-Feb-2014
Datum / Date: