
	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ3D00.01
---	---	-----------

TECHNOVIEW Apixaban CONT High 5x1ml

REF 5090270
LOT 3D41B00.01
 29-Feb-2016

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.4 – 7.8	7.5	OK
Virologie Virology	HBsAg / HBsAg HIV-AK / HIV-Ab HCV-AK / HCV-Ab	Negative Negative Negative	Negative Negative Negative	OK
Chargenspezifische Kalibrationswerte Lot specific calibration values		2.36 – 3.93U/mL	3.14U/mL	OK
		261.2 – 435.4ng/mL	348.3ng/mL	
Verwendbarkeit Usability Ceveron ® alpha	Anti-Xa Apixaban 3931B00	261.2 – 435.4ng/mL	349.7	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

Durchführung von
Prepared by


Lucas HOLOVICZ

28-Feb-2014
Datum / Date:

Freigabe von
Approved by


Alessandro BENIGNI

28-Feb-2014
Datum / Date: