



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQ3D00.01

TECHNOVIEW Apixaban CONT High 5x1mL

REF 5090270

LOT 3D41B00.02

29-Feb-2016

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.4 – 7.8	7.5	OK
Virologie Virology	HBsAg / HBsAg HIV-AK / HIV-Ab HCV-AK / HCV-Ab	Negative Negative Negative	Negative Negative Negative	OK
Chargenspezifische Kalibrationswerte Lot specific calibration values		2.36 – 3.93U/mL	3.14U/mL	OK
		261.2 – 435.4ng/mL	348.3ng/mL	
Verwendbarkeit Usability Ceveron ® alpha	Anti-Xa Apixaban LOT 3D41B00.02	261.2 – 435.4ng/mL	353.9	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

Durchführung von
Prepared by

Lucas ROSENMAYR

12-Mär-2015
Datum / Date:

Freigabe von
Approved by

Lavdrim NASUFI

12-Mär-2015
Datum / Date: