



| | | |
|---|---|-----------|
|  | ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS | DQ3D00.01 |
|---|---|-----------|


TECHNOVIEW Apixaban CONT High 5x1mL

REF 5090270
LOT 3D41B01.01
 31-Jul-2017


| Analytische Prozedur Analytical Procedure | | Anforderung Requirement | Ergebnis Result | |
|---|--|----------------------------------|----------------------------------|----|
| pH | | 7.4 – 7.8 | 7.5 | OK |
| Virologie Virology | HBsAg / HBsAg HIV-AK / HIV-Ab HCV-AK / HCV-Ab | Negative Negative Negative | Negative Negative Negative | OK |
| Chargenspezifische Kalibrationswerte Lot specific calibration values | | 2.36 – 3.93U/mL | 3.14U/mL | OK |
| | | 261.2 – 435.4ng/mL | 348.3ng/mL | |
| Verwendbarkeit Usability Ceveron ® alpha | Anti-Xa Apixaban LOT 3D41B01.01 | 261.2 – 435.4ng/mL | 396.2 | OK |
| Etiketten und Barcode Überprüfung Label and bar code Verification | | | OK | |

**Durchführung von
Prepared by**


Lucas ROSENMAYR


Datum / Date:

**Freigabe von
Approved by**


Lucas HOLOVICZ


Datum / Date: