



ANALYSENZERTIFIKAT  
CERTIFICATE OF ANALYSIS

DQ3D00.01

TECHNOVIEW Apixaban CONT High 5x1mL

REF 5090270

LOT 3D43B01.01

⌚ 31-Jul-2017

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.4 – 7.8	7.6	OK
Virologie Virology	HBsAg / HBsAg HIV-AK / HIV-Ab HCV-AK / HCV-Ab	Negative Negative Negative	Negative Negative Negative	OK
Chargenspezifische Kalibrationswerte Lot specific calibration values		2.42 – 4.04 IU/mL	3.23	OK
		282.3 – 470.5 ng/mL	376.4	
Verwendbarkeit Usability Ceveron ® alpha	Anti-Xa Apixaban	282.3 – 470.5 ng/mL	380.1	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

Durchführung von  
Prepared by


  
Lucas ROSENMAYR

21. DEZ. 2015

Datum / Date:

Freigabe von  
Approved by

  
Michael KAFKA

  
Datum / Date: