
	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ3900.01
---	---	-----------

TECHNOVIEW Apixaban CONT Low 5x1ml

REF 5090271
LOT 3932B00.01
 31-Dec-2015

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.4 – 7.8	7.6	OK
Virologie Virology	HBsAg / HBsAg HIV-AK / HIV-Ab HCV-AK / HCV-Ab	Negative Negative Negative	Negative Negative Negative	OK
Chargenspezifische Kalibrationswerte Lot specific calibration values		0.65 – 1.09U/mL	0.87	OK
		98.9 – 164.8ng/mL	131.8	
Verwendbarkeit Usability Ceveron © alpha	Anti-Xa Apixaban 3932B00	98.9 – 164.8ng/mL	122.0	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

**Durchführung von
Prepared by**

Lucas HOLOVICZ

25-Feb-2014
Datum / Date:

**Freigabe von
Approved by**

Lavdrim NASUFI

25-Feb-2014
Datum / Date: