
	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ3900.01
---	---	-----------

TECHNOVIEW Apixaban CONT Low 5x1ml

REF 5090271
LOT 3941B00.02
 29-Feb-2016

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.4 – 7.8	7.7	OK
Virologie Virology	HBsAg / HBsAg HIV-AK / HIV-Ab HCV-AK / HCV-Ab	Negative Negative Negative	Negative Negative Negative	 OK
Chargenspezifische Kalibrationswerte Lot specific calibration values		0.71 – 1.19U/mL	0.95	OK
		104.2 – 173.6ng/mL	138.9	
Verwendbarkeit Usability Ceveron ® alpha	Anti-Xa Apixaban 3932B00	104.2 – 173.6ng/mL	157.7	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

Durchführung von
Prepared by

Lucas HOLOVICZ

26-Jun-2014
Datum / Date:

Freigabe von
Approved by

Alessandro BENIGNI

26-Jun-2014
Datum / Date: