



ANALYSENZERTIFIKAT  
CERTIFICATE OF ANALYSIS

DQ3900.01

TECHNOVIEW Apixaban CONT Low 5x1mL

REF 5090271

LOT 3941B00.03

⌚ 29-Feb-2016

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.4 – 7.8	7.7	OK
Virologie Virology	HBsAg / HBsAg HIV-AK / HIV-Ab HCV-AK / HCV-Ab	Negative Negative Negative	Negative Negative Negative	OK
Chargenspezifische Kalibrationswerte Lot specific calibration values		0.71 – 1.19U/mL	0.95	OK
		104.2 – 173.6ng/mL	138.9	
Verwendbarkeit Usability Ceveron ® alpha	Anti-Xa Apixaban LOT 3941B00.03	104.2 – 173.6ng/mL	143.3	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

Durchführung von  
Prepared by

  
Lucas ROSENMAYR

12-Feb-2015  
Datum / Date:

Freigabe von  
Approved by

  
Lavdrim NASUFI

12-Mar-2015  
Datum / Date: