
	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ3900.01
---	---	-----------

TECHNOVIEW Apixaban CONT Low 5x1ml

REF 5090271
LOT 3941B01.01
 31-Jul-2017


Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.4 – 7.8	7.7	OK
Virologie Virology	HBsAg / HBsAg HIV-AK / HIV-Ab HCV-AK / HCV-Ab	Negative Negative Negative	Negative Negative Negative	OK
Chargenspezifische Kalibrationswerte Lot specific calibration values		0.71 – 1.19 IU/mL	0.95	OK
		104.2 – 173.6 ng/mL	138.9	
Verwendbarkeit Usability Ceveron ® alpha	Anti-Xa Apixaban	104.2 – 173.6 ng/mL	132.5	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

Durchführung von
Prepared by


Lucas ROSENMAYR

27-Aug-2015
Datum / Date:

Freigabe von
Approved by


Monika WEINBAUER

27-Aug-2015
Datum / Date: