



ANALYSENZERTIFIKAT  
CERTIFICATE OF ANALYSIS

DQ3900.01

TECHNOVIEW Apixaban CONT Low 5x1ml

REF 5090271

LOT 3943B01.01

🕒 31-Jul-2017

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.4 – 7.8	7.6	OK
Virologie Virology	HBsAg / HBsAg HIV-AK / HIV-Ab HCV-AK / HCV-Ab	Negative Negative Negative	Negative Negative Negative	OK
Chargenspezifische Kalibrationswerte Lot specific calibration values		0.81 – 1.51 IU/mL	1.16	OK
		114.1 – 211.9 ng/mL	163.0	
Verwendbarkeit Usability Ceveron® alpha	Anti-Xa Apixaban	114.1 – 211.9 ng/mL	167.7	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

Durchführung von  
Prepared by

  
Lucas ROSENMAYR

21. DEZ. 2015  
Datum / Date:

Freigabe von  
Approved by

  
Michael KAFKA

21-DEC-2015  
Datum / Date: