



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQ6E0AAZ.01


Fibrinogen Reagent 5x2mL

REF 5138085
LOT 6E52BA0.01
⌚ 31-Mar-2017

Analytische Prozedur Analytical Procedure	Anforderung Requirement	Ergebnis Result	
pH	7.0 – 8.0	7.8	OK
Verwendbarkeit Usability Ceveron [®] alpha	CCN, LOT 1P37B00	2.96	OK
	CCA, LOT 3P38B00	0.93 – 1.39 g/L	0.99
Etiketten und Barcode Überprüfung Label and bar code Verification		OK	


Durchführung von
Prepared by


Monika WEINBAUER


Datum / Date:

Freigabe von
Approved by


Fikret NASUFI


Datum / Date: