
	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ470DAZ.01
---	---	-------------

Factor VII deficient plasma, depleted 5x1mL

REF 5144015
LOT 4741B0D.01
 30-Jun-2016

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.4 – 7.8	7.7	OK
Virologie Virology	HBsAg / HBsAg HIV-AK / HIV-Ab HCV-AK / HCV-Ab	Negative Negative Negative	Negative Negative Negative	OK
Verwendbarkeit Usability Ceveron® alpha	CCN LOT 1P44B00	87.8 – 146.3 %	101.8	OK
	CCA LOT 3P11B00	31.5 – 52.5 %	42.1	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

Durchführung von
Prepared by

Alessandro BENIGNI

29-Sep-2016
Datum / Date:

Freigabe von
Approved by

Lucas HOLOVICZ

29-Sep-2016
Datum / Date: