

	<b>ANALYSENZERTIFIKAT</b> <b>CERTIFICATE OF ANALYSIS</b>	DQ0F00AZ.02
---	---	-------------

### F VIII Inhibitor Reagent Kit 2-4T

REF      5152005  
LOT      0F41B00.01  
      30-Apr-2015

Inhalt Content	Reagenz Reagent	<span style="border: 1px solid black; padding: 2px;">LOT</span>	
2x2.29mL	FVIII normal Plasma	1G31B00	30-Jun-2015
1x1mL	FVIII Inhibitor Plasma	2K31B00	30-Nov-2015
1x1mL	FVIII Inhibitor Free Plasma	2M31B0D	30-Apr-2015
1x17mL	Imidazol Puffer	9J32BB0	30-Jun-2018

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Verwendbarkeit Usability	FVIII Inhibitor Plasma (2K31B00)	9.5 – 15.9 BE/mL	13.1	OK
FVIII clotting, Daptin TC Ceveron® alpha	FVIII Inhibitor Free Plasma (2M31B0D)	< 0.4 BE/mL	< 0.4	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

**Durchführung von**  
**Prepared by**

Alessandro BENIGNI

17-Jan-2014  
Datum / Date:

**Freigabe von**  
**Approved by**

Lucas HOLOVICZ

17-Jan-2014  
Datum / Date: