



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQ0F00AZ.02

F VIII Inhibitor Reagent Kit 2-4T

REF 5152005
LOT 0F41B00.02
⌚ 30-Apr-2015

Inhalt Content	Reagenz Reagent	LOT	⌚
2x2.29mL	FVIII normal Plasma	1G31B00	30-Jun-2015
1x1mL	FVIII Inhibitor Plasma	2K31B00	30-Nov-2015
1x1mL	FVIII Inhibitor Free Plasma	2M31B0D	30-Apr-2015
1x17mL	Imidazol Puffer	9J32BB0	30-Jun-2018

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Verwendbarkeit Usability FVIII clotting, Daptin TC Ceveron [®] alpha	FVIII Inhibitor Plasma LOT 2K31B00	9.5 – 15.9 BE/mL	13.6	OK
	FVIII Inhibitor Free Plasma LOT 2M31B0D	< 0.4 BE/mL	< 0.4	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

Durchführung von
Prepared by

Alessandro BENIGNI

16-Apr-2014
Datum / Date:

Freigabe von
Approved by

Lucas HOŁOVICZ

16-Apr-2014
Datum / Date: