



ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS

DQ0F00AZ.02

F VIII Inhibitor Reagent Kit 2-4T

REF 5152005
LOT 0F53B00.01
⌚ 31-Aug-2016

Inhalt Content	Reagenz Reagent	LOT	⌚
2x2.29mL	FVIII normal Plasma	1G31B01	31-Aug-2016
1x1mL	FVIII Inhibitor Plasma	2K31B01	30-Apr-2017
1x1mL	FVIII Inhibitor Free Plasma	2M41B0D	31-Oct-2016
1x17mL	Imidazol Puffer	9J49BA0	31-Oct-2019

Analytische Prozedur Analytical Procedure	Anforderung Requirement	Ergebnis Result
Verwendbarkeit Usability FVIII clotting, Daptin TC Ceveron [®] alpha	FVIII Inhibitor Plasma LOT 2K31B01 9.5 – 15.9 BE/mL	13.4 OK
	FVIII Inhibitor Free Plasma LOT 2M41B0D < 0.4 BE/mL	< 0.4 OK
Etiketten und Barcode Überprüfung Label and bar code Verification		OK

Durchführung von
Prepared by

Lucas ROSENMAYR

06-Aug-2015
Datum / Date:

Freigabe von
Approved by

Fikret NASUFI

06-Aug-2015
Datum / Date: