

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ480DAZ.02
---	---	-------------

Factor VIII deficient plasma, depleted 5x1mL

REF 5154002
LOT 4841B0D.01
 31-Jan-2016


Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.4 – 7.8	7.8	OK
Virologie Virology	HBsAg / HBsAg HIV-AK / HIV-Ab HCV-AK / HCV-Ab	Negative Negative Negative	Negative Negative Negative	OK
Restfeuchte Residual moisture		≤1%	0.04	OK
Verwendbarkeit Usability Ceveron® alpha	CCN, LOT 1P14B00	72 – 120 %	101.0	OK
	CCA, LOT 3P24B00	12.0 – 20.0 %	15.1	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

Durchführung von
Prepared by


Lavdrim NASUFI

01-Apr-2014
Datum / Date:

Freigabe von
Approved by


Lucas HOLOVICZ

07-Apr-2014
Datum / Date: