

	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ480D.01
---	---	-----------

F VIII Mangelplasma, immunads. 5x1mL

REF 5154002
LOT 4851B0D.01
 31-Mar-2017

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.4 – 7.8	7.5	OK
Virologie Virology	HBsAg / HBsAg HIV-AK / HIV-Ab HCV-AK / HCV-Ab	Negative Negative Negative	Negative Negative Negative	OK
Restfeuchte Residual moister		≤1%	0.34	OK
Verwendbarkeit Usability Ceveron [®] alpha	CCN LOT 1P37B00	65.3 – 108.8 %	101.6	OK
	CCA LOT 3P38B00	15.0 – 25.0 %	20.2	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

Durchführung von
Prepared by


Lavdrim NASUFI

28-Apr-2015
Datum / Date:

Freigabe von
Approved by


Alessandro BENIGNI

28-Apr-2015
Datum / Date: