

	<b>ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS</b>	DQ4800AZ.01
---	---	-------------

**Factor VIII deficient plasma, nativ 5x1mL**

REF      5154007  
LOT      4861B00.01  
      31-Jan-2018

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.4 – 7.8	7.8	OK
Virologie Virology	HBsAg / HBsAg HIV-AK / HIV-Ab HCV-AK / HCV-Ab	Negative Negative Positive	Negative Negative Positive	OK
Verwendbarkeit Usability Ceveron® alpha	CCN, <span style="border: 1px solid black; padding: 1px;">LOT</span> 1P37B00	65.3 – 108.8 %	77.8	OK
	CCA, <span style="border: 1px solid black; padding: 1px;">LOT</span> 3P38B00	15.0 – 25.0 %	21.2	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

**Durchführung von  
Prepared by**

  
**Monika WEINBAUER**

29. FEB. 2016

Datum / Date:

**Freigabe von  
Approved by**

  
**Lucas ROSENMAYR**

29. FEB. 2016

Datum / Date: