
	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ480DAZ.02
---	---	-------------


Factor VIII deficient plasma, native. 50x1mL


REF 5154016
LOT 4831B00.01
 28-Feb-2015

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.4 – 7.8	7.5	OK
Virologie Virology	HBsAg / HBsAg HIV-AK / HIV-Ab HCV-AK / HCV-Ab	Negative Negative Positive	Negative Negative Positive	OK
Verwendbarkeit Usability Ceveron® alpha	CCN (1P14B00)	72 – 120 %	92.5 %	OK
	CCA (3P24B00)	11.9 – 19.9 %	16.0 %	OK

Durchführung von
Prepared by

Alessandro BENIGNI




Datum / Date:

Freigabe von
Approved by

Harald SEVCIK



04. JULI 2013

Datum / Date: