



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQ4800AZ.02


Factor VIII deficient plasma, native 50x1mL

REF 5154016
LOT 4842B00.01
⌚ 29-Feb-2016

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Fibrinogengehalt Fibrinogen content	Clauss method	>1.5g/L	2.1	OK
Restfeuchte Residual moister		≤1%	0.04	OK
OD _{320nm}		≤10	5.52	OK
Verwendbarkeit Usability Ceveron® alpha	CCN (1P14B00)	72.0 – 120.0 %	109.0	OK
	CCA (3P24B00)	12.0 – 20.0 %	15.5	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

Durchführung von
Prepared by


Lavdrim NASUFI



06-Mär-2014
Datum / Date:

Freigabe von
Approved by

Lucas HOLOVICZ



06-Mär-2014
Datum / Date: