



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQ4800AZ.02

Factor VIII deficient plasma, nativ 50x1mL

REF 5154016

LOT 4845B00.01

31-Aug-2016

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.4 – 7.8	7.8	OK
Virologie Virology	HBsAg / HBsAg HIV-AK / HIV-Ab HCV-AK / HCV-Ab	Negative Negative Positive	Negative Negative Positive	OK
Verwendbarkeit Usability Ceveron [®] alpha	CCN, LOT 1P44B00	69.0 – 115.0 %	80.2	OK
	CCA, LOT 3P38B00	15.0 – 25.0 %	21.1	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

Durchführung von
Prepared by

Alessandro BENIGNI

04-Nov-2014
Datum / Date:

Freigabe von
Approved by

Lucas HOLOVIEZ

04-Nov-2014
Datum / Date: