
	ANALYSENZERTIFIKAT CERTIFICATE OF ANALYSIS	DQ490DAZ.02
---	---	-------------

Factor IX deficient plasma, immunads. 5x1mL

REF 5164003
LOT 4931B0F.01
 31-May-2017

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.4 – 7.8	7.5	OK
Virologie Virology	HBsAg / HBsAg HIV-AK / HIV-Ab HCV-AK / HCV-Ab	Negative Negative Negative	Negative Negative Negative	OK
Verwendbarkeit Usability Ceveron® alpha	CCN LOT 1P37B00	80.3 – 133.8 %	121.0	OK
	CCA LOT 3P38B00	15.8 – 26.3 %	21.4	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

Durchführung von
Prepared by


Lucas ROSENMAYR



17-Jul-2015
Datum / Date:

Freigabe von
Approved by

Lucas HOLOVICZ



17-Jul-2015
Datum / Date: