



ANALYSENZERTIFIKAT
CERTIFICATE OF ANALYSIS

DQ490DAZ.02

Factor IX deficient plasma, nativ 5x1mL

REF 5164008
LOT 4923B01.01
⌚ 31-Aug-2016

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.4 – 7.8	7.6	OK
Virologie Virology	HBsAg / HBsAg HIV-AK / HIV-Ab HCV-AK / HCV-Ab	Negative Negative Positive	Negative Negative Positive	OK
Verwendbarkeit Usability Ceveron [®] alpha	CCN LOT 1P44B00	94.5 – 157.5 %	131.2 %	OK
	CCA LOT 3P11B00	17.0 – 28.0 %	19.9 %	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

Durchführung von
Prepared by

Alessandro BENIGNI

29-Sep-2014

Datum / Date:

Freigabe von
Approved by

Lucas HOLOVICZ

29-Sep-2014

Datum / Date: