



ANALYSENZERTIFIKAT  
CERTIFICATE OF ANALYSIS

DQ4000AZ.01

Factor X deficient plasma nativ, 5x1mL

REF 5174004

LOT 4022B02.01

⌚ 31-Jul-2017

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
pH		7.4 – 7.8	7.7	OK
Virologie Virology	HBsAg / HBsAg HIV-AK / HIV-Ab HCV-AK / HCV-Ab	Negative Negative Positive	Negative Negative Positive	OK
Verwendbarkeit Usability Ceveron® alpha	CCN LOT 1P37B00	78 – 130 %	94.2	OK
	CCA LOT 3P38B00	15 – 25 %	18.4	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

Durchführung von  
Prepared by

Fikret NASUFI  
*Fikret Nasufi*

10-Sep-2015  
Datum / Date:

Freigabe von  
Approved by

Alessandro BENIGNI  
*Alessandro Benigni*

10-Sep-2015  
Datum / Date: