



ANALYSENZERTIFIKAT  
CERTIFICATE OF ANALYSIS

DQ4Z00AZ.02

Factor XII Mangelplasma, nativ 5x1mL

REF 5194008

LOT 4Z41B00.01

28-Feb-2017

Analytische Prozedur Analytical Procedure		Anforderung Requirement	Ergebnis Result	
Fibrinogengehalt Fibrinogen content		> 1.5g/L	1.93	OK
Virologie Virology	HBsAg / HBsAg HIV-AK / HIV-Ab HCV-AK / HCV-Ab	Negative Negative Positive	Negative Negative Positive	OK
Verwendbarkeit Usability Ceveron® alpha	CCN [LOT] 1P44B00	81.8 – 136.3 %	101.3	OK
	CCA [LOT] 3P24B00	29.3 – 48.8 %	40.0	OK
Etiketten und Barcode Überprüfung Label and bar code Verification			OK	

Durchführung von  
Prepared by

Lavdrim NASUFI

20-Jun-2014  
Datum / Date:

Freigabe von  
Approved by

Lucas HOLOVICZ

20-Jun-2014  
Datum / Date: